



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

MP COMPONENTS
1805 Lemon Farris Road
Cookeville, TN 38506
John Schlatter Phone: 931 525 1600

MECHANICAL

Valid To: October 31, 2022

Certificate Number: 2247.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following dimensional testing/calibrations:

I. Dimensional Testing/Calibration¹

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
3D Measurements ³ – Steel Aluminum Non-Metal	Up to (40 × 60 × 35) in	(550 + 43L) μin (560 + 54L) μin (560 + 110L) μin	Bridge CMM (Chameleon 1298-2490)
Steel Aluminum Non-Metal	Up to (78 × 130 × 60) in	(930 + 36L) μin (930 + 47L) μin (940 + 100L) μin	Gantry CMM (Alpha 018)
Length ³ – Outside Diameter	Up to 4 in	0.0001 in	Micrometer

¹ This laboratory offers commercial dimensional testing/calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

³This laboratory meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.

⁴In the statement of CMC, *L* represents the numerical value of the nominal length of the device measured in inches.



Accredited Laboratory

A2LA has accredited

MP COMPONENTS

Cookeville, TN

for technical competence in the field of

Mechanical Testing

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 7th day of July 2020

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2247.02
Valid to October 31, 2022

For the types of tests to which this accreditation applies, please refer to the laboratory's Mechanical Scope of Accreditation.