



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

MP COMPONENTS INC.  
8499 Centre Industrial Drive  
Byron Center, MI 49315  
Randy Veldhouse Phone: 616 878 9710

MECHANICAL

Valid To: October 31, 2026

Certificate Number: 2247.01

In recognition of the successful completion of the A2LA evaluation process accreditation is granted to this laboratory to perform the following dimensional testing<sup>1, 5</sup>:

I. Dimensional Testing<sup>1</sup>

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
3D Measurements <sup>3</sup> – Steel Aluminum Non-Metal	Up to (40 × 60 × 35) in	(390 + 46L) μin (400 + 57L) μin (410 + 120L) μin	Bridge CMM (Scirocco 022)
Steel Aluminum Non-Metal	Up to (78 × 130 × 60) in	(790 + 39L) μin (790 + 50L) μin (790 + 110L) μin	Gantry CMM (Alpha 042)
Steel Aluminum Non-Metal	Up to (40 × 60 × 35) in	(420 + 45L) μin (420 + 57L) μin (440 + 120L) μin	Bridge CMM (Xcel 1294-608)
Length <sup>3</sup> – Outside Diameter	Up to 3 in	0.0001 in	Micrometers

<sup>1</sup> This laboratory offers commercial dimensional testing service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC uncertainty due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> This test is not equivalent to that of a calibration.

<sup>4</sup> In the statement of CMC,  $L$  represents the numerical value of the nominal length of the device measured in inches.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## Accredited Laboratory

A2LA has accredited

### MP COMPONENTS INC.

*Byron Center, MI*

for technical competence in the field of

### Mechanical Testing

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing laboratories*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 15<sup>th</sup> day of October 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2247.01  
Valid to October 31, 2026

*For the types of tests to which this accreditation applies, please refer to the laboratory's Mechanical Scope of Accreditation.*